

The 7060 and 7068 water retracting relays function the same, switching water on/off with an air signal. The difference is that the 7060 when shutting off sucks back about four inches of water depending on tubing size. This is a issue when the handpiece is in a place where is can suck back blood or saliva. The 7068 will not suck back any water but this means that the water will be right at the tip of the handpiece and may drip out.



**7060**  
**Retracting**



**7068**  
**Non-Retracting**